

PRE-BID QUERIES/CLARIFICATIONS (Annexure-A)
(I-ITN21001 Dated 26 April,2021)

Name of the Bidder	Query No.	Ref. Tender Part/Section No.	Ref. Clause No.	Description of the Query	ITER-India Response
Bidder-1	1	Optical Fiber bundle Assembly	3 (Table no.1, Sr. No.1)	Termination-Connector B &C Required OD 35.0 mm- We offer OD-48.10 mm(Shell Size 25/MS Sheel Size Code J)	Fiber bundle connector Outer Diameter (OD) requirements is updated as per Amendment
	2	Optical Fiber bundle Assembly	3 (Table no.1, Sr. No.6)	Overall Transmission Efficiency including coupling requirement is 80-90% (i.e. dB)- We offer approx. 50% i.e. insertion loss of connector 2.5 dB for each fiber line	The overall transmission Efficiency including bundle coupling requirement should be>80%.
	3	Optical Fiber bundle Assembly	3 (Table no.1, Sr. No.7)	Fiber transmission attenuation: We offer fiber with attenuation 19.32 dB/km (468nm) 12.45dB/km (525nm) and 5.87 dB/km (656nm)	Required fiber transmission attenuation is 30-50 dB/Km & offered (5-20dB/Km) is better than the required and help to achieve the overall better transmission so it is acceptable
	4	Essential Eligibility Criteria (EEC)	Table No. 1	Requirement- in-place certified quality management system as per ISO 9001:2015, else should have its own documented QMS- we are ISO 13485 quality management system and comply with all QMS documents. The details on this can be shared if required. Prior to ISO 13485 we were ISO 9001 compliant for more than 3 years	The bidder should have a certification if not, they should share the QMS as defined in the Essential Eligibility Criteria (EEC)